
Midmark M11 - v1.0.0
Total Cycles: 3233

Name: RITTER 1
Sterilizer ID: V2178198
05/15/2023 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.0 F 0.0
2:00 186.0 F 0.0
4:00 200.0 F 4.0
6:00 217.8 F 9.0
8:00 226.1 F 9.7
10:00 233.6 F 9.8
12:00 238.2 F 8.7
14:00 244.3 F 13.4
16:00 243.6 F 12.3
18:00 256.8 F 19.1
20:00 268.2 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.2
2:00 271.6 F 28.4
3:00 272.9 F 29.3
4:00 272.3 F 28.8

Min 271.5 F 28.4
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 211.0 F 0.0
10:00 236.1 F 0.0
15:00 230.4 F 0.0
20:00 228.7 F 0.0
25:00 228.6 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 21:14
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:25
CYCLE COMPLETE

05/15/2023 02:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____