

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1978

Name: RITTER2  
Sterilizer ID: V2313228  
01/30/2023 12:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.8 F 0.0  
2:00 163.6 F 3.3  
4:00 179.2 F 7.2  
6:00 200.9 F 12.5  
8:00 214.7 F 7.1  
10:00 235.0 F 12.4  
12:00 235.9 F 9.7  
14:00 236.2 F 8.9  
16:00 249.3 F 15.3  
18:00 261.3 F 22.4  
20:00 272.3 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.6  
1:00 272.7 F 29.9  
2:00 272.5 F 30.3  
3:00 272.4 F 29.6  
4:00 272.9 F 30.3

Min 271.8 F 29.5  
Max 273.0 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.6 F 0.0  
5:00 222.8 F 0.0  
10:00 219.1 F 0.0  
15:00 216.4 F 0.0  
20:00 231.3 F 0.0  
25:00 235.8 F 0.0  
30:00 222.3 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 21:17  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:46  
CYCLE COMPLETE

01/30/2023 01:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_