
Midmark M11 - v1.0.5
Total Cycles: 1777

Name: RITTER2
Sterilizer ID: V2313228
11/28/2022 12:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.6 F 0.0
2:00 193.8 F 3.7
4:00 209.4 F 9.4
6:00 226.6 F 8.2
8:00 236.2 F 12.1
10:00 238.4 F 10.7
12:00 238.3 F 10.1
14:00 252.8 F 17.4
16:00 265.0 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.4 F 29.5
2:00 272.6 F 30.1
3:00 272.0 F 29.2
4:00 273.0 F 30.0

Min 271.7 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 227.3 F 0.0
10:00 234.5 F 0.0
15:00 232.1 F 0.0
20:00 221.0 F 0.0
25:00 217.3 F 0.0
30:00 213.1 F 0.0
Drying Complete

FILLING: 2:40
HEATING: 17:42
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:33
CYCLE COMPLETE

11/28/2022 01:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____