
Midmark M11 - v1.0.0
Total Cycles: 4415

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2024 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 73.8 F 0.1
4:00 103.4 F 1.0
6:00 118.7 F 3.1
8:00 138.5 F 5.8
10:00 166.0 F 9.0
12:00 184.5 F 12.4
14:00 202.0 F 2.5
16:00 212.8 F 5.6
18:00 222.2 F 8.9
20:00 232.6 F 12.6
22:00 232.2 F 8.2
24:00 243.5 F 13.2
26:00 241.6 F 11.2
28:00 253.6 F 17.2
30:00 264.2 F 23.5
32:00 272.0 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.0
2:00 271.5 F 28.3
3:00 272.7 F 29.1
4:00 271.8 F 28.4

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 210.5 F 0.0
10:00 212.9 F 0.0
15:00 214.7 F 0.0
20:00 216.1 F 0.0
25:00 219.1 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 32:10
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:10:29
CYCLE COMPLETE

03/28/2024 10:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____