
Midmark M11 - v1.0.5
Total Cycles: 516

Name: RITTER2
Sterilizer ID: V2313228
11/03/2021 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 81.2 F 0.0
2:00 84.5 F 0.0
4:00 91.0 F 0.0
6:00 109.7 F 0.2
8:00 155.7 F 0.3
10:00 182.3 F 2.0
12:00 197.8 F 5.3
14:00 209.9 F 8.6
16:00 221.1 F 12.4
18:00 226.6 F 7.9
20:00 237.0 F 12.5
22:00 237.4 F 10.3
24:00 237.0 F 8.9
26:00 248.9 F 15.1
28:00 260.0 F 21.4
30:00 270.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.5 F 29.5
2:00 271.7 F 29.1
3:00 272.6 F 29.6
4:00 271.8 F 28.9

Min 271.6 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 208.0 F 0.0
10:00 214.1 F 0.0
15:00 234.1 F 0.0
20:00 228.2 F 0.0
25:00 225.9 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 30:51
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:02
CYCLE COMPLETE

11/03/2021 12:06 PM
ACCEPTED: X
REJECTED:
Operator: U01