
Midmark M11 - v1.0.0
Total Cycles: 4071

Name: RITTER 1
Sterilizer ID: V2178198
11/30/2023 01:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.8 F 0.0
2:00 116.8 F 1.1
4:00 128.6 F 3.6
6:00 157.0 F 6.8
8:00 186.0 F 10.2
10:00 208.9 F 4.1
12:00 215.2 F 6.5
14:00 225.7 F 10.5
16:00 229.9 F 7.6
18:00 238.6 F 11.4
20:00 236.4 F 9.2
22:00 249.5 F 15.3
24:00 260.8 F 21.5
26:00 271.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.1 F 28.9
2:00 272.4 F 29.3
3:00 271.7 F 28.6
4:00 272.5 F 29.2

Min 271.5 F 28.5
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 213.5 F 0.0
10:00 223.2 F 0.0
15:00 239.3 F 0.0
20:00 233.2 F 0.0
25:00 228.3 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 26:41
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:05:13
CYCLE COMPLETE

11/30/2023 02:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____