
Midmark M11 - v1.0.5
Total Cycles: 269

Name: RITTER2
Sterilizer ID: V2313228
08/09/2021 08:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.6 F 0.0
2:00 108.6 F 0.7
4:00 112.0 F 2.5
6:00 125.8 F 5.7
8:00 157.9 F 9.8
10:00 181.7 F 6.8
12:00 210.4 F 6.9
14:00 223.7 F 11.5
16:00 233.3 F 8.6
18:00 238.1 F 11.3
20:00 244.8 F 13.1
22:00 256.5 F 19.3
24:00 266.7 F 25.7
26:00 271.5 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 272.6 F 29.9
2:00 272.1 F 30.0
3:00 272.5 F 29.8
4:00 272.0 F 30.0

Min 271.6 F 29.5
Max 273.3 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.0 F 0.0
5:00 210.3 F 0.0
10:00 223.9 F 0.0
15:00 182.1 F 0.0
20:00 185.5 F 0.0
25:00 181.7 F 0.0
30:00 177.6 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:35
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:04:43
CYCLE COMPLETE

08/09/2021 09:37 AM
ACCEPTED: X
REJECTED:
Operator: U01