
Midmark M11 - v1.0.0
Total Cycles: 3915

Name: RITTER 1
Sterilizer ID: V2178198
10/17/2023 04:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.2 F 0.0
2:00 184.9 F 3.9
4:00 196.5 F 8.6
6:00 219.6 F 8.1
8:00 230.9 F 9.3
10:00 237.9 F 9.1
12:00 243.8 F 12.9
14:00 243.2 F 11.9
16:00 257.1 F 19.1
18:00 269.2 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.4 F 28.8
2:00 271.4 F 28.4
3:00 272.5 F 28.9
4:00 271.5 F 28.2

Min 271.3 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 213.5 F 0.0
10:00 236.1 F 0.0
15:00 229.7 F 0.0
20:00 225.9 F 0.0
25:00 219.1 F 0.0
30:00 213.4 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 18:60
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

10/17/2023 05:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____