

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3445

Name: RITTER2  
Sterilizer ID: V2313228  
03/08/2024 11:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.1 F 0.0  
2:00 135.6 F 0.4  
4:00 141.5 F 2.2  
6:00 166.3 F 5.4  
8:00 187.3 F 9.2  
10:00 205.9 F 12.3  
12:00 220.2 F 6.9  
14:00 229.9 F 11.6  
16:00 233.6 F 8.6  
18:00 240.5 F 11.5  
20:00 244.9 F 13.1  
22:00 256.2 F 19.2  
24:00 266.7 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.9  
1:00 272.1 F 29.6  
2:00 272.1 F 29.3  
3:00 272.8 F 29.9  
4:00 271.6 F 29.3

Min 270.9 F 28.7  
Max 273.2 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.7 F 0.0  
5:00 213.2 F 0.0  
10:00 203.7 F 0.0  
15:00 199.1 F 0.0  
20:00 195.9 F 0.0  
25:00 193.1 F 0.0  
30:00 191.4 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 25:41  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:04:11  
CYCLE COMPLETE

03/08/2024 12:44 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_