
Midmark M11 - v1.0.5
Total Cycles: 1250

Name: RITTER2
Sterilizer ID: V2313228
07/04/2022 01:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.0 F 0.0
2:00 117.7 F 1.4
4:00 141.8 F 4.2
6:00 170.1 F 8.0
8:00 191.1 F 12.6
10:00 211.4 F 5.8
12:00 223.9 F 10.6
14:00 228.6 F 7.3
16:00 239.6 F 12.3
18:00 239.2 F 10.7
20:00 251.8 F 17.0
22:00 263.1 F 23.6
24:00 272.4 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.6 F 29.8
2:00 271.3 F 29.1
3:00 272.7 F 30.0
4:00 271.9 F 29.4

Min 271.2 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.5 F 0.0
5:00 201.6 F 0.0
10:00 215.9 F 0.0
15:00 217.8 F 0.0
20:00 214.3 F 0.0
25:00 211.1 F 0.0
30:00 208.4 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 24:14
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:02:27
CYCLE COMPLETE

07/04/2022 02:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____