
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002563
9:11 07/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:02
STERILIZING 3:07
136.5 C 222kPa 3:07
136.4 C 225kPa 3:37
136.3 C 223kPa 4:07
136.1 C 225kPa 4:37
136.4 C 224kPa 5:07
136.2 C 222kPa 5:37
136.4 C 225kPa 6:07
136.4 C 222kPa 6:37
136.4 C 222kPa 6:37
VENTING 6:37
AIR DRYING 7:32
CYCLE COMPLETE 37:32

Digital Signature #
B26DACFFE0A91236
