
Midmark M11 - v1.0.5
Total Cycles: 2964

Name: RITTER2
Sterilizer ID: V2313228
10/04/2023 07:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 79.6 F 0.0
4:00 104.9 F 0.0
6:00 112.9 F 0.0
8:00 152.2 F 0.0
10:00 182.5 F 0.0
12:00 206.0 F 0.0
14:00 211.9 F 0.0
16:00 215.8 F 1.2
18:00 226.6 F 4.8
20:00 235.6 F 8.4
22:00 245.1 F 12.7
24:00 241.0 F 10.8
26:00 237.7 F 9.4
28:00 238.4 F 9.2
30:00 251.7 F 16.1
32:00 264.2 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.1 F 28.4
2:00 272.7 F 28.8
3:00 272.9 F 29.2
4:00 271.9 F 28.6

Min 271.7 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.3 F 0.0
5:00 228.9 F 0.0
10:00 196.9 F 0.0
15:00 192.9 F 0.0
20:00 190.7 F 0.0
25:00 191.6 F 0.0
30:00 200.6 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 33:51
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:11:45
CYCLE COMPLETE

10/04/2023 09:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____