
Midmark M11 - v1.0.5
Total Cycles: 1729

Name: RITTER2
Sterilizer ID: V2313228
11/15/2022 03:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.2 F 0.0
2:00 204.1 F 3.7
4:00 210.4 F 9.9
6:00 225.1 F 5.9
8:00 241.1 F 12.9
10:00 241.4 F 11.6
12:00 241.6 F 11.1
14:00 256.2 F 18.9
16:00 269.0 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 272.5 F 28.8
2:00 273.0 F 29.3
3:00 272.1 F 29.0
4:00 272.5 F 28.9

Min 271.8 F 28.5
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 227.8 F 0.0
10:00 223.8 F 0.0
15:00 230.2 F 0.0
20:00 231.3 F 0.0
25:00 229.5 F 0.0
30:00 227.1 F 0.0

Drying Complete

FILLING: 2:50
HEATING: 16:59
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 00:56:45
CYCLE COMPLETE

11/15/2022 04:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____