

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001218  
16:35 17/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:43  
STERILIZING 3:52  
136.5 C 225kPa 3:52  
136.3 C 224kPa 4:23  
136.4 C 226kPa 4:53  
136.3 C 225kPa 5:23  
136.4 C 225kPa 5:53  
136.4 C 223kPa 6:23  
136.2 C 223kPa 6:53  
136.2 C 224kPa 7:23  
136.2 C 224kPa 7:23  
VENTING 7:23  
AIR DRYING 8:12  
CYCLE COMPLETE 12:34  
-----

Digital Signature #  
23A7071119911CE1  
-----