
Midmark M11 - v1.0.5
Total Cycles: 2357

Name: RITTER2
Sterilizer ID: V2313228
05/11/2023 01:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.5 F 0.0
2:00 195.8 F 4.1
4:00 207.6 F 9.6
6:00 224.7 F 7.0
8:00 234.7 F 11.7
10:00 236.8 F 9.9
12:00 236.8 F 8.9
14:00 250.6 F 16.1
16:00 262.5 F 23.1
18:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.5 F 29.7
2:00 271.6 F 29.0
3:00 272.7 F 29.9
4:00 272.0 F 29.2

Min 271.5 F 29.0
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.7 F 0.0
5:00 213.1 F 0.0
10:00 215.1 F 0.0
15:00 201.8 F 0.0
20:00 199.9 F 0.0
25:00 201.6 F 0.0
30:00 208.8 F 0.0

Drying Complete

FILLING: 2:40
HEATING: 18:14
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:13
CYCLE COMPLETE

05/11/2023 02:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____