

-----  
Midmark M11 - v1.0.5  
Total Cycles: 496

Name: RITTER2  
Sterilizer ID: V2313228  
10/28/2021 09:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.1 F 0.0  
2:00 168.8 F 3.4  
4:00 187.9 F 8.0  
6:00 206.2 F 13.3  
8:00 220.6 F 7.2  
10:00 233.5 F 12.7  
12:00 236.6 F 10.2  
14:00 236.2 F 8.7  
16:00 249.8 F 15.7  
18:00 261.8 F 22.6  
20:00 272.7 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.6 F 29.6  
2:00 271.4 F 28.9  
3:00 272.6 F 29.7  
4:00 271.8 F 29.0

Min 271.4 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.5 F 0.0  
5:00 234.8 F 0.0  
10:00 227.0 F 0.0  
15:00 217.9 F 0.0  
20:00 215.9 F 0.0  
25:00 214.1 F 0.0  
30:00 212.5 F 0.0  
Drying Complete

FILLING: 1:35  
HEATING: 20:21  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 00:58:58  
CYCLE COMPLETE

10/28/2021 10:49 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02