

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4460

Name: RITTER2  
Sterilizer ID: V2313228  
03/26/2025 01:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.4 F 0.0  
2:00 159.9 F 2.5  
4:00 169.1 F 6.2  
6:00 192.1 F 10.8  
8:00 220.3 F 4.6  
10:00 223.6 F 9.4  
12:00 236.3 F 10.4  
14:00 238.9 F 11.7  
16:00 238.6 F 10.3  
18:00 250.9 F 16.6  
20:00 262.2 F 23.0  
22:00 272.7 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.7  
1:00 272.3 F 30.4  
2:00 271.5 F 29.8  
3:00 271.9 F 30.2  
4:00 271.6 F 30.3

Min 270.9 F 29.6  
Max 273.7 F 31.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.7 F 0.0  
5:00 240.3 F 0.0  
10:00 218.4 F 0.0  
15:00 209.3 F 0.0  
20:00 193.4 F 0.0  
25:00 191.8 F 0.0  
30:00 191.8 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 23:13  
STERILIZING: 4:00  
VENTING: 3:53  
DRYING: 30:00  
TOTAL CYCLE: 01:02:25  
CYCLE COMPLETE

03/26/2025 02:26 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_