
Midmark M11 - v1.0.5
Total Cycles: 2042

Name: RITTER2
Sterilizer ID: V2313228
02/13/2023 02:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.7 F 0.0
2:00 107.4 F 0.0
4:00 125.9 F 0.0
6:00 160.3 F 0.8
8:00 184.2 F 3.6
10:00 207.4 F 7.3
12:00 213.5 F 11.5
14:00 224.1 F 6.7
16:00 236.1 F 11.8
18:00 235.6 F 9.5
20:00 237.9 F 9.7
22:00 248.2 F 14.9
24:00 260.0 F 21.5
26:00 270.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.5 F 29.6
2:00 271.6 F 29.5
3:00 272.4 F 29.6
4:00 271.4 F 29.3

Min 271.4 F 29.1
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.7 F 0.0
5:00 222.4 F 0.0
10:00 219.4 F 0.0
15:00 197.2 F 0.0
20:00 196.1 F 0.0
25:00 195.8 F 0.0
30:00 201.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:46
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:04:59
CYCLE COMPLETE

02/13/2023 03:51 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____