

-----  
Midmark M11 - v1.0.5  
Total Cycles: 649

Name: RITTER2  
Sterilizer ID: V2313228  
12/20/2021 08:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.2 F 0.0  
2:00 85.0 F 0.0  
4:00 83.7 F 0.1  
6:00 148.4 F 2.5  
8:00 183.0 F 5.7  
10:00 209.2 F 9.6  
12:00 214.4 F 8.7  
14:00 223.1 F 7.5  
16:00 235.0 F 12.0  
18:00 232.9 F 8.2  
20:00 242.8 F 12.5  
22:00 242.0 F 11.8  
24:00 254.4 F 18.2  
26:00 265.6 F 25.2  
28:00 273.3 F 30.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.7  
1:00 271.7 F 29.8  
2:00 271.8 F 29.8  
3:00 271.8 F 29.8  
4:00 271.8 F 29.7

Min 271.6 F 29.5  
Max 273.2 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.4 F 0.0  
5:00 240.0 F 0.0  
10:00 181.5 F 0.0  
15:00 206.0 F 0.0  
20:00 202.6 F 0.0  
25:00 214.0 F 0.0  
Dry Cycle Aborted

FILLING: 0:50  
HEATING: 28:16  
STERILIZING: 4:00  
VENTING: 2:52  
DRYING: 26:28  
TOTAL CYCLE: 01:02:26

12/20/2021 09:03 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02