

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001830  
12:58 21/06/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:15  
PRESSURIZING 3:19  
STERILIZING 4:45  
136.5 C 220kPa 4:45  
136.1 C 218kPa 5:15  
136.4 C 224kPa 5:45  
136.3 C 223kPa 6:15  
136.3 C 223kPa 6:45  
136.4 C 224kPa 7:15  
136.3 C 222kPa 7:45  
136.3 C 224kPa 8:15  
136.3 C 224kPa 8:15  
VENTING 8:15  
AIR DRYING 9:09  
CYCLE COMPLETE 9:10  
-----

Digital Signature #  
48CB7F636B118DCD  
-----