
Midmark M11 - v1.0.0
Total Cycles: 3831

Name: RITTER 1
Sterilizer ID: V2178198
09/28/2023 07:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.3 F 0.0
2:00 81.6 F 0.0
4:00 84.1 F 0.0
6:00 124.6 F 0.0
8:00 142.9 F 0.3
10:00 172.1 F 2.4
12:00 195.7 F 4.6
14:00 221.8 F 8.0
16:00 226.8 F 11.6
18:00 225.3 F 5.5
20:00 235.9 F 10.3
22:00 236.0 F 8.5
24:00 244.1 F 12.3
26:00 241.1 F 10.5
28:00 253.9 F 17.0
30:00 265.0 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.0 F 28.1
2:00 272.5 F 28.5
3:00 271.8 F 28.4
4:00 271.9 F 27.9

Min 271.4 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.6 F 0.0
5:00 221.4 F 0.0
10:00 220.6 F 0.0
15:00 215.1 F 0.0
20:00 228.6 F 0.0
25:00 225.2 F 0.0
30:00 221.5 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:48
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:09:48
CYCLE COMPLETE

09/28/2023 08:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____