
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001042
7:16 29/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:19
PRESSURIZING 3:20
STERILIZING 4:21
136.0 C 220kPa 4:21
136.0 C 220kPa 4:51
136.4 C 222kPa 5:21
135.8 C 219kPa 5:51
136.1 C 222kPa 6:21
136.1 C 220kPa 6:51
136.3 C 222kPa 7:21
136.4 C 224kPa 7:51
136.4 C 224kPa 7:51
VENTING 7:52
AIR DRYING 8:42
CYCLE COMPLETE 8:56

Digital Signature #
F65DF9F6CC69548A
