
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007297
7:17 02/05/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:33
STERILIZING 3:41
136.4 C 220kPa 3:41
135.9 C 220kPa 4:11
136.4 C 224kPa 4:41
136.4 C 223kPa 5:11
136.4 C 221kPa 5:41
136.0 C 221kPa 6:11
136.1 C 218kPa 6:41
136.4 C 222kPa 7:11
136.4 C 222kPa 7:11
VENTING 7:11
AIR DRYING 8:00
CYCLE COMPLETE 15:15

Digital Signature #
967D750E8E612506
