
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000367
9:19 31/08/2020

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:50
STERILIZING 3:58
136.4 C 222kPa 3:58
136.4 C 224kPa 4:28
136.4 C 222kPa 4:58
136.4 C 220kPa 5:28
136.1 C 221kPa 5:58
136.4 C 222kPa 6:28
136.2 C 221kPa 6:58
136.2 C 220kPa 7:28
136.2 C 220kPa 7:28
VENTING 7:28
AIR DRYING 8:19
CYCLE COMPLETE 8:55

Digital Signature #
2AA4FB2962FCC6D9
