
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001612
12:18 20/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:07
STERILIZING 4:14
136.5 C 226kPa 4:14
136.2 C 223kPa 4:44
136.4 C 226kPa 5:14
136.1 C 223kPa 5:44
136.4 C 226kPa 6:14
136.4 C 227kPa 6:44
136.4 C 226kPa 7:14
136.4 C 226kPa 7:44
136.4 C 226kPa 7:44
VENTING 7:44
AIR DRYING 8:39
CYCLE COMPLETE 13:18

Digital Signature #
6077189B6B9B5AE1
