
Midmark M11 - v1.0.0
Total Cycles: 1102

Name: RITTER 1
Sterilizer ID: V2178198
08/19/2021 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.6 F 0.0
2:00 166.5 F 3.6
4:00 184.5 F 7.8
6:00 221.3 F 13.2
8:00 224.3 F 7.5
10:00 238.5 F 12.7
12:00 238.8 F 10.0
14:00 236.0 F 8.2
16:00 250.6 F 15.4
18:00 263.1 F 22.6
20:00 271.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.6 F 28.7
2:00 271.3 F 28.2
3:00 272.8 F 28.8
4:00 271.6 F 28.2

Min 271.2 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.9 F 0.0
5:00 205.1 F 0.0
10:00 225.6 F 0.0
15:00 233.3 F 0.0
20:00 226.1 F 0.0
25:00 222.6 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 20:04
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

08/19/2021 03:00 PM
ACCEPTED: X
REJECTED:
Operator: U02