

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5439

Name: RITTER 1  
Sterilizer ID: V2178198  
03/24/2025 05:54 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.9 F 0.0  
2:00 162.5 F 3.7  
4:00 186.3 F 8.3  
6:00 213.4 F 12.5  
8:00 224.4 F 7.8  
10:00 237.5 F 13.0  
12:00 237.2 F 10.0  
14:00 235.9 F 8.2  
16:00 249.0 F 15.0  
18:00 261.2 F 21.9  
20:00 272.4 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.4 F 29.2  
2:00 271.4 F 28.5  
3:00 272.5 F 29.3  
4:00 271.8 F 28.6  
  
Min 271.3 F 28.4  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.0 F 0.0  
5:00 199.1 F 0.0  
10:00 199.2 F 0.0  
15:00 210.6 F 0.0  
20:00 231.2 F 0.0  
25:00 237.7 F 0.0  
30:00 230.8 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 20:27  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 00:59:19  
CYCLE COMPLETE

03/24/2025 06:53 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_