
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000835
15:31 31/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:20
CYCLE ABORTED 6:05
WATER LEVEL LOW
NOT STERILE
END TIME 6:05

Digital Signature #
CBCA550D74B32D1A
