

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2820

Name: RITTER2  
Sterilizer ID: V2313228  
08/30/2023 03:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 146.3 F 0.0  
2:00 143.6 F 0.0  
4:00 172.6 F 0.0  
6:00 201.9 F 3.0  
8:00 220.6 F 7.2  
10:00 230.6 F 11.8  
12:00 231.5 F 8.0  
14:00 243.0 F 13.5  
16:00 242.4 F 11.9  
18:00 241.0 F 11.1  
20:00 254.4 F 17.9  
22:00 266.0 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.3 F 29.1  
2:00 272.1 F 29.4  
3:00 272.0 F 28.8  
4:00 272.5 F 29.5

Min 271.2 F 28.6  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.9 F 0.0  
5:00 231.3 F 0.0  
10:00 212.4 F 0.0  
15:00 195.1 F 0.0  
20:00 185.4 F 0.0  
25:00 184.4 F 0.0  
30:00 182.9 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 23:37  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:01:55  
CYCLE COMPLETE

08/30/2023 04:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_