

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4274

Name: RITTER 1  
Sterilizer ID: V2178198  
02/06/2024 10:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.3 F 0.0  
2:00 146.2 F 1.4  
4:00 146.3 F 3.9  
6:00 162.7 F 7.2  
8:00 185.3 F 11.1  
10:00 212.3 F 3.5  
12:00 219.1 F 7.7  
14:00 231.0 F 12.0  
16:00 232.3 F 8.2  
18:00 243.8 F 13.4  
20:00 242.1 F 11.4  
22:00 253.3 F 17.1  
24:00 263.6 F 23.2  
26:00 272.6 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.8  
1:00 272.8 F 29.5  
2:00 272.0 F 28.8  
3:00 272.7 F 29.5  
4:00 272.2 F 29.0

Min 271.6 F 28.7  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.3 F 0.0  
5:00 227.1 F 0.0  
10:00 230.8 F 0.0  
15:00 225.6 F 0.0  
20:00 234.4 F 0.0  
25:00 229.9 F 0.0  
30:00 230.0 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 26:57  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:05:37  
CYCLE COMPLETE

02/06/2024 11:50 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_