

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2425

Name: RITTER 1  
Sterilizer ID: V2178198  
10/13/2022 07:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.8 F 0.0  
2:00 77.2 F 0.0  
4:00 81.6 F 0.1  
6:00 103.4 F 0.6  
8:00 135.7 F 3.2  
10:00 170.7 F 6.4  
12:00 192.4 F 10.5  
14:00 220.2 F 5.3  
16:00 221.6 F 6.7  
18:00 231.0 F 10.5  
20:00 235.0 F 8.6  
22:00 239.4 F 11.3  
24:00 235.2 F 8.5  
26:00 248.7 F 14.7  
28:00 259.4 F 20.8  
30:00 269.4 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.2 F 29.0  
2:00 272.7 F 29.4  
3:00 272.1 F 29.4  
4:00 271.7 F 28.9

Min 271.4 F 28.8  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.3 F 0.0  
5:00 221.6 F 0.0  
10:00 223.5 F 0.0  
15:00 210.9 F 0.0  
20:00 203.9 F 0.0  
25:00 205.8 F 0.0  
30:00 199.2 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 31:40  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:09:44  
CYCLE COMPLETE

10/13/2022 08:10 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_