

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001200  
9:10 18/02/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:00  
PRESSURIZING 3:02  
STERILIZING 4:02  
136.4 C 220kPa 4:02  
136.4 C 225kPa 4:32  
136.4 C 224kPa 5:02  
136.4 C 221kPa 5:32  
136.2 C 223kPa 6:02  
136.1 C 219kPa 6:32  
135.9 C 220kPa 7:02  
136.4 C 222kPa 7:32  
136.4 C 222kPa 7:32  
VENTING 7:33  
AIR DRYING 8:24  
CYCLE COMPLETE 8:34

-----  
Digital Signature #  
7B2DA0C1E547240B  
-----