
Midmark M11 - v1.0.0
Total Cycles: 2170

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2022 01:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.6 F 0.0
2:00 194.3 F 4.5
4:00 200.0 F 9.5
6:00 221.0 F 6.0
8:00 230.8 F 10.6
10:00 232.4 F 7.4
12:00 243.6 F 12.3
14:00 244.3 F 12.6
16:00 256.9 F 19.3
18:00 268.3 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 29.1
2:00 271.3 F 28.4
3:00 272.4 F 29.3
4:00 272.0 F 28.7

Min 271.3 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 238.7 F 0.0
10:00 230.0 F 0.0
15:00 228.1 F 0.0
20:00 228.7 F 0.0
25:00 230.9 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

07/25/2022 02:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____