
Midmark M11 - v1.0.5
Total Cycles: 3998

Name: RITTER2
Sterilizer ID: V2313228
10/03/2024 06:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.4 F 0.0
2:00 139.4 F 2.4
4:00 165.1 F 6.0
6:00 194.8 F 11.0
8:00 210.2 F 4.5
10:00 224.5 F 9.4
12:00 233.1 F 9.6
14:00 239.6 F 11.8
16:00 239.4 F 10.6
18:00 253.0 F 17.4
20:00 264.6 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.7 F 29.5
2:00 271.8 F 29.3
3:00 272.5 F 29.4
4:00 271.7 F 29.2

Min 271.6 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 199.1 F 0.0
10:00 205.2 F 0.0
15:00 211.3 F 0.0
20:00 213.4 F 0.0
25:00 215.8 F 0.0
30:00 214.6 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 21:54
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:00:36
CYCLE COMPLETE

10/03/2024 07:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____