
Midmark M11 - v1.0.0
Total Cycles: 3042

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2023 07:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 80.3 F 0.0
2:00 80.7 F 0.0
4:00 92.1 F 0.0
6:00 126.6 F 0.0
8:00 162.7 F 1.9
10:00 174.9 F 5.5
12:00 192.9 F 9.2
14:00 207.9 F 13.0
16:00 219.1 F 6.0
18:00 230.2 F 9.9
20:00 235.4 F 9.9
22:00 239.8 F 11.0
24:00 235.8 F 8.5
26:00 249.7 F 15.0
28:00 260.9 F 21.3
30:00 271.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.9
2:00 271.4 F 28.1
3:00 272.7 F 29.0
4:00 271.8 F 28.2

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.3 F 0.0
5:00 210.3 F 0.0
10:00 217.9 F 0.0
15:00 222.6 F 0.0
20:00 220.2 F 0.0
25:00 226.8 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:41
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:41
CYCLE COMPLETE

03/28/2023 08:40 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____