

-----  
Midmark M11 - v1.0.5  
Total Cycles: 136

Name: RITTER2  
Sterilizer ID: V2313228  
06/24/2021 08:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.1 F 0.0  
2:00 75.2 F 0.2  
4:00 79.8 F 1.0  
6:00 96.2 F 2.9  
8:00 123.8 F 5.6  
10:00 155.3 F 9.0  
12:00 180.1 F 12.9  
14:00 209.6 F 5.0  
16:00 216.8 F 8.6  
18:00 231.1 F 12.4  
20:00 231.5 F 8.6  
22:00 241.1 F 13.5  
24:00 242.4 F 11.6  
26:00 253.6 F 17.2  
28:00 263.4 F 23.1  
30:00 272.6 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 271.5 F 29.4  
2:00 271.6 F 29.4  
3:00 271.4 F 29.4  
4:00 271.6 F 29.3

Min 271.3 F 29.1  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.0 F 0.0  
5:00 242.8 F 0.0  
10:00 234.6 F 0.0  
15:00 219.1 F 0.0  
20:00 197.4 F 0.0  
25:00 198.8 F 0.0  
30:00 197.8 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 30:56  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:08:57  
CYCLE COMPLETE

06/24/2021 09:47 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_