

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001371  
11:07 30/11/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:46  
PRESSURIZING 4:47  
STERILIZING 6:12  
136.4 C 225kPa 6:12  
136.3 C 224kPa 6:42  
136.4 C 221kPa 7:12  
136.3 C 224kPa 7:42  
136.4 C 224kPa 8:12  
136.3 C 223kPa 8:42  
136.3 C 223kPa 9:12  
136.2 C 221kPa 9:42  
136.2 C 221kPa 9:42  
VENTING 9:42  
AIR DRYING 10:41  
CYCLE COMPLETE 25:41  
-----

Digital Signature #  
A9D19B8AB6B674C8  
-----