
Midmark M11 - v1.0.5
Total Cycles: 3723

Name: RITTER2
Sterilizer ID: V2313228
06/20/2024 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.9 F 0.0
2:00 127.9 F 1.3
4:00 144.1 F 3.8
6:00 169.9 F 7.5
8:00 190.3 F 11.9
10:00 221.2 F 5.6
12:00 226.5 F 10.3
14:00 231.9 F 7.5
16:00 241.4 F 12.0
18:00 238.7 F 10.2
20:00 251.8 F 16.7
22:00 263.4 F 23.5
24:00 272.1 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.0
1:00 272.5 F 29.7
2:00 271.3 F 28.9
3:00 272.3 F 29.9
4:00 273.1 F 30.0

Min 271.0 F 28.9
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 244.0 F 0.0
10:00 204.2 F 0.0
15:00 204.5 F 0.0
20:00 218.9 F 0.0
25:00 228.5 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 24:12
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:02:32
CYCLE COMPLETE

06/20/2024 11:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____