
Midmark M11 - v1.0.0
Total Cycles: 4859

Name: RITTER 1
Sterilizer ID: V2178198
08/28/2024 11:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.0 F 0.0
2:00 158.6 F 0.3
4:00 178.6 F 1.0
6:00 201.8 F 5.2
8:00 215.7 F 10.2
10:00 223.2 F 6.2
12:00 234.3 F 10.0
14:00 234.9 F 8.3
16:00 246.2 F 13.5
18:00 244.9 F 12.4
20:00 257.9 F 19.4
22:00 269.5 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.3 F 28.3
2:00 272.5 F 28.7
3:00 271.7 F 27.9
4:00 272.7 F 28.7

Min 271.5 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.5 F 0.0
5:00 207.4 F 0.0
10:00 207.2 F 0.0
15:00 212.5 F 0.0
20:00 230.8 F 0.0
25:00 227.6 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 22:57
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:01:08
CYCLE COMPLETE

08/28/2024 12:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____