

-----  
Midmark M11 - v1.0.0  
Total Cycles: 928

Name: RITTER 1  
Sterilizer ID: V2178198  
06/23/2021 09:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.2 F 0.0  
2:00 172.1 F 1.8  
4:00 179.5 F 4.8  
6:00 200.1 F 8.9  
8:00 215.6 F 13.4  
10:00 225.2 F 7.3  
12:00 235.3 F 11.8  
14:00 234.2 F 8.6  
16:00 241.5 F 11.1  
18:00 244.9 F 12.7  
20:00 256.4 F 18.8  
22:00 266.6 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.5 F 29.1  
2:00 271.6 F 28.4  
3:00 272.1 F 29.1  
4:00 272.2 F 28.8

Min 271.3 F 28.3  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.5 F 0.0  
5:00 218.8 F 0.0  
10:00 218.7 F 0.0  
15:00 209.9 F 0.0  
20:00 217.7 F 0.0  
25:00 220.7 F 0.0  
30:00 225.2 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 23:39  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:02:41  
CYCLE COMPLETE

06/23/2021 10:23 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_