
Midmark M11 - v1.0.0
Total Cycles: 3617

Name: RITTER 1
Sterilizer ID: V2178198
08/04/2023 10:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.5 F 0.0
2:00 135.8 F 0.0
4:00 144.0 F 0.0
6:00 177.6 F 3.0
8:00 192.5 F 6.6
10:00 204.7 F 10.3
12:00 219.1 F 8.3
14:00 225.5 F 7.4
16:00 235.5 F 11.6
18:00 233.3 F 8.1
20:00 245.0 F 13.3
22:00 243.2 F 11.8
24:00 255.0 F 17.9
26:00 265.6 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.9
2:00 271.7 F 28.1
3:00 272.0 F 28.7
4:00 272.4 F 28.7

Min 271.4 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 209.4 F 0.0
10:00 207.6 F 0.0
15:00 218.9 F 0.0
20:00 213.2 F 0.0
25:00 212.0 F 0.0
30:00 210.5 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 27:52
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:06:16
CYCLE COMPLETE

08/04/2023 11:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____