
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 005063
11:18 03/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:44
STERILIZING 3:29
136.0 C 220kPa 3:29
136.4 C 223kPa 3:59
136.3 C 220kPa 4:29
136.3 C 222kPa 4:59
136.1 C 219kPa 5:29
136.3 C 222kPa 5:59
136.4 C 223kPa 6:29
136.2 C 220kPa 6:59
136.2 C 220kPa 6:59
VENTING 6:59
AIR DRYING 7:41
CYCLE COMPLETE 37:41

Digital Signature #
F502CFFE8AF960A2
