

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2544

Name: RITTER2  
Sterilizer ID: V2313228  
06/21/2023 09:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.5 F 0.0  
2:00 192.6 F 3.4  
4:00 194.4 F 8.3  
6:00 213.1 F 10.2  
8:00 228.1 F 9.0  
10:00 235.9 F 10.0  
12:00 242.9 F 12.8  
14:00 242.6 F 11.9  
16:00 255.7 F 18.7  
18:00 267.0 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.1 F 29.5  
2:00 272.7 F 29.4  
3:00 272.0 F 29.0  
4:00 271.4 F 28.6  
  
Min 271.1 F 28.5  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.8 F 0.0  
5:00 221.4 F 0.0  
10:00 234.8 F 0.0  
15:00 234.0 F 0.0  
20:00 217.2 F 0.0  
25:00 215.4 F 0.0  
30:00 216.4 F 0.0  
  
Drying Complete

FILLING: 1:50  
HEATING: 19:34  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:36  
CYCLE COMPLETE

06/21/2023 10:34 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_