

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1464

Name: RITTER 1  
Sterilizer ID: V2178198  
12/16/2021 07:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.5 F 0.0  
2:00 82.1 F 0.0  
4:00 91.8 F 0.0  
6:00 125.2 F 0.2  
8:00 148.0 F 1.2  
10:00 176.0 F 3.3  
12:00 203.6 F 6.2  
14:00 216.0 F 9.5  
16:00 228.4 F 13.3  
18:00 229.2 F 7.6  
20:00 239.5 F 12.1  
22:00 234.6 F 8.6  
24:00 242.0 F 11.3  
26:00 244.6 F 12.6  
28:00 256.3 F 18.8  
30:00 267.0 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.7 F 29.1  
2:00 271.5 F 28.3  
3:00 272.7 F 29.2  
4:00 272.1 F 28.7

Min 271.5 F 28.3  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.7 F 0.0  
5:00 212.4 F 0.0  
10:00 210.4 F 0.0  
15:00 174.3 F 0.0  
20:00 192.8 F 0.0  
25:00 203.6 F 0.0  
30:00 206.7 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 31:32  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 01:09:33  
CYCLE COMPLETE

12/16/2021 09:41 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02