

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002134  
7:33 08/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:35  
PRESSURIZING 2:44  
STERILIZING 3:52  
136.0 C 220kPa 3:52  
136.2 C 220kPa 4:22  
136.4 C 224kPa 4:52  
136.0 C 222kPa 5:22  
136.3 C 224kPa 5:52  
136.4 C 225kPa 6:22  
136.3 C 224kPa 6:52  
136.0 C 221kPa 7:22  
136.0 C 221kPa 7:22  
VENTING 7:22  
AIR DRYING 8:14  
CYCLE COMPLETE 8:37  
-----

Digital Signature #  
EC52C2B401CB370F  
-----