
Midmark M11 - v1.0.0
Total Cycles: 4009

Name: RITTER 1
Sterilizer ID: V2178198
11/13/2023 09:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.7 F 0.0
2:00 69.3 F 0.1
4:00 77.3 F 0.9
6:00 114.6 F 2.8
8:00 148.1 F 5.3
10:00 169.1 F 7.9
12:00 187.8 F 11.0
14:00 215.7 F 5.1
16:00 218.7 F 5.3
18:00 230.3 F 8.5
20:00 237.4 F 12.1
22:00 230.7 F 7.5
24:00 242.3 F 12.3
26:00 239.1 F 9.7
28:00 250.3 F 15.3
30:00 260.8 F 21.1
32:00 270.1 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 271.3 F 28.6
2:00 271.4 F 28.5
3:00 271.9 F 28.6
4:00 272.4 F 28.8

Min 271.2 F 28.5
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.0 F 0.0
5:00 247.6 F 0.0
10:00 228.5 F 0.0
15:00 220.9 F 0.0
20:00 194.2 F 0.0
25:00 188.3 F 0.0
30:00 184.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 33:28
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:11:58
CYCLE COMPLETE

11/13/2023 10:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____