
Midmark M11 - v1.0.5
Total Cycles: 1723

Name: RITTER2
Sterilizer ID: V2313228
11/15/2022 09:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.4 F 0.0
2:00 176.5 F 0.6
4:00 191.1 F 4.5
6:00 211.7 F 10.3
8:00 222.2 F 6.5
10:00 237.6 F 12.3
12:00 238.2 F 10.6
14:00 237.9 F 9.6
16:00 252.2 F 16.8
18:00 264.2 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.9
1:00 272.8 F 29.6
2:00 271.8 F 29.1
3:00 272.5 F 29.3
4:00 271.9 F 29.2

Min 271.7 F 28.8
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 209.9 F 0.0
10:00 220.5 F 0.0
15:00 225.5 F 0.0
20:00 228.2 F 0.0
25:00 224.3 F 0.0
30:00 210.9 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 19:49
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:10
CYCLE COMPLETE

11/15/2022 09:59 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____