

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001806  
10:44 26/06/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:05  
PRESSURIZING 2:11  
STERILIZING 3:15  
136.0 C 221kPa 3:15  
136.0 C 222kPa 3:45  
136.2 C 224kPa 4:15  
136.3 C 224kPa 4:45  
136.2 C 222kPa 5:15  
136.2 C 223kPa 5:45  
136.4 C 226kPa 6:15  
136.4 C 224kPa 6:45  
136.4 C 224kPa 6:45  
VENTING 6:45  
AIR DRYING 7:36  
CYCLE COMPLETE 31:56  
-----

Digital Signature #  
9A21A2FAF9095E92  
-----