
Midmark M11 - v1.0.0
Total Cycles: 969

Name: RITTER 1
Sterilizer ID: V2178198
07/06/2021 12:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.9 F 0.0
2:00 130.2 F 0.8
4:00 148.1 F 3.0
6:00 176.6 F 6.6
8:00 199.9 F 10.6
10:00 217.4 F 5.0
12:00 228.0 F 8.9
14:00 236.5 F 11.5
16:00 238.8 F 10.8
18:00 235.5 F 8.6
20:00 249.9 F 15.2
22:00 261.2 F 21.7
24:00 271.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.5 F 29.1
2:00 271.4 F 28.3
3:00 272.6 F 29.2
4:00 271.9 F 28.6

Min 271.2 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.0 F 0.0
5:00 231.1 F 0.0
10:00 219.2 F 0.0
15:00 223.3 F 0.0
20:00 226.6 F 0.0
25:00 230.8 F 0.0
30:00 237.7 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 24:36
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:02:55
CYCLE COMPLETE

07/06/2021 01:59 PM
ACCEPTED: X
REJECTED:
Operator: U01