
Midmark M11 - v1.0.0
Total Cycles: 3058

Name: RITTER 1
Sterilizer ID: V2178198
03/30/2023 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.7 F 0.0
2:00 198.7 F 4.6
4:00 202.4 F 9.5
6:00 217.3 F 6.4
8:00 230.4 F 9.8
10:00 233.7 F 8.3
12:00 244.2 F 13.0
14:00 243.3 F 11.9
16:00 256.5 F 18.9
18:00 268.3 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.0
2:00 271.7 F 28.3
3:00 272.7 F 29.2
4:00 272.4 F 28.8

Min 271.5 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 218.0 F 0.0
10:00 233.8 F 0.0
15:00 228.8 F 0.0
20:00 226.8 F 0.0
25:00 226.4 F 0.0
30:00 225.4 F 0.0

Drying Complete

FILLING: 1:40
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:11
CYCLE COMPLETE

03/30/2023 11:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____